

CALIBRATION CERTIFICATE

NAME OF COMPANY : AL MUHALIB READY MIX CO
Address : RIYADH
Location : RIYADH
Date of Issue : 28-10-2025
Cert. Ref. No. : 2K25103003 - 11

IDENTIFICATION ITEM UNDER CALIBRATION : **COMPRESSION TEST MACHINE**
Capacity: 2000kn x 0.1kn
Serial No: 24/000608
Make/Model: UTEST, UTC-5727.FPR


SUMMARY OF MEASUREMENT RESULT : The above compression machine has been calibrated by using our SASO Standard Load cell with indicator and found within tolerance limit.

DATE OF CALIBRATION : 28-10-2025

CALIBRATION DUE ON : 27-04-2026


CALIBRATED BY
HOM




APPROVED BY
HOM

CALIBRATION REPORT

NAME OF CO.	AL MUHALIB READY MIX CO
ADDRESS	RIYADH
LOCATION	RIYADH

SCALE TYPE	COMPRESSION TEST MACHINE	SERIAL NO.	24/000608
MODEL	UTEST	MODEL #	UTEST, UTC-5727.FPR
CAPACITY	2000 KN	DATE OF CAL.	28-10-2025
PER DIVISION	0.1KN	CERT REF #	2K25103003 - 11

The above compression machine has been calibrated by using our Standard Load cell with indicator and found within tolerance limit. The result of the measurement is as follows:

APPLIED LOAD (KN)	TEST 1	TEST 2	TEST 3	MEAN (KN)	ERROR % OF RDG.	ACCURACY OF EQUIP. %
0	0.00	0.00	0.00	0.00	0.00	0.00 %
10	10.00	10.00	10.10	10.04	0.40	99.60 %
20	19.91	20.10	20.10	20.04	0.20	99.80 %
30	30.60	30.40	29.42	30.14	0.47	99.53 %
50	49.43	50.41	50.90	50.24	0.48	99.52 %
100	99.00	99.93	102.19	100.37	0.37	99.63 %
200	195.84	202.12	205.45	201.13	0.57	99.44 %
300	302.83	302.54	299.10	301.49	0.50	99.50 %
400	411.78	402.07	393.25	402.37	0.59	99.41 %
500	501.12	504.55	503.08	502.92	0.58	99.42 %
600	598.99	598.21	611.94	603.04	0.51	99.49 %
800	795.71	797.77	815.13	802.87	0.36	99.64 %
1,000	1007.73	1005.18	1004.40	1005.77	0.58	99.42 %
1,500	1512.68	1512.19	1501.89	1508.92	0.59	99.41 %
2,000	2016.44	2022.13	1991.93	2010.17	0.51	99.49 %

Accuracy	Grade 1.0	Compression	10 KN	Down To	2000 KN
Repeatability	Grade 1.0	Compression	10 KN	Down To	2000 KN

[Signature]

HOM



[Signature]
House of Measurement